

TECHNICAL SPECIFICATION

Procurement of a Jaw Crusher

Approved by

Production Director
Underground Mining Operations

LLC "RMG Auramine"

 V. Abashmadze
" 09 " 06 2026

Agreed by

Chief Geologist

LLC "RMG Auramine"

 N. Tatuashvili
" 09 " 06 2026

1. Procurement Object

This technical specification defines the minimum technical requirements for a jaw crusher intended for the Geological Sample Preparation Laboratory of the Bertakari Deposit operated by LLC "RMG Auramine".

2. Scope of Application

The jaw crusher shall be intended for the primary crushing of geological samples, rocks, and ore materials.

3. Scope of Supply

The supply shall include:

1. Jaw crusher;
2. All major components required for the operation of the equipment.

4. Technical Requirements

No.	Parameter	Requirement
1	Equipment Type	Jaw Crusher
2	Capacity	Average 2.5–3 tons/hour
3	Maximum Feed Size	170–200 mm
4	Output Size	6–30 mm
5	Discharge Gap Adjustment	Required
6	Construction	Heavy-duty industrial design
7	Jaw Material	Wear-resistant steel or equivalent
8	Power Supply	380V / 50Hz / Three-phase
9	Safety System	Required
10	Warranty	Minimum 12 months
11	Compliance with Standards	CE Certificate or equivalent conformity document
12	Technical Documentation	Technical passport, operating instructions, and certificates
13	Condition	New, unused, factory-manufactured equipment

Prepared by

Head of Sample Preparation Laboratory

LLC "RMG Auramine"

5.3 მმ Avtandil Vardzelashvili

"09" 06 2026