

## დანართი 1- ტექნიკური მახასიათებლები და რაოდენობა

დასახელება: Covid 19-ის ანტიგენის სწრაფი ტესტი

რაოდენობა: 1410 ცალი

ტესტი უნდა აკმაყოფილებდეს შემდეგ მოთხოვნებს:

1. ტესტი განკუთვნილი უნდა იყოს თვითგამოყენებისთვის, არ მოითხოვდეს სპეციალურად გადამზადებული პერსონალის ჩარევას.
2. ტესტების ინდივიდუალურად ან მცირე რაოდენობით (2-5-მდე) უნდა იყოს შეფუთული (ე.ი. ცალკეული ტესტების ბუფერული ხსნარი უნდა იყოს ხელმისაწვდომი, მაგ. სინჯარაში)
3. ნაცხი ცხვირის ნესტოში (არა ნერწყვი, სისხლი, ნაზოფარინგეალური ან ხახის ნაცხი)
4. აღიარებული უნდა იყოს მსოფლიო ჯანდაცვის ორგანიზაციისა ან ევროკავშირის (ან მისი რომელიმე წევრი ქვეყანის) მიერ.
5. ვარგისიანობა უნდა ჰქონდეს მინიმუმ 2022 წლის ბოლომდე
6. ტესტის პასუხი 10-20 წუთი

English version

### **Annex 1- Technical specifications and quantities**

Description: Covid 19 Rapid Antigen Test

Quantity - 1410

The test must meet the following requirements:

1. The test shall be intended for self-use and shall not require the intervention of specially trained personnel.
2. Tests should be packaged individually or in small quantities (up to 2-5) (a buffer solution for individual tests should be available, in a test tube)
3. Smear in nasal swab (no saliva, blood, nasopharyngeal or throat swab)
4. Must be approved either by WHO or by EU (or it's member states)
5. Must be valid min. till the end of 2022
6. Test should yield a result in a reasonable time (10-20 minutes)